

## CLARIFICATION NO. 1

TENDER ENQUIRY NO. PROC-FE/CB/DRG-STIM/4319/2019

DESCRIPTION: CEMENT LAB EQUIPMENTS

**Item # -01- HPHT Dual Cell Consistometer**

**Query:** Kindly confirm which of the following is required:

| Sr # | Description  | OGDCL Remarks |
|------|--|---------------|
| 1    | Automatic Pressure Control   | Required      |
| 2    | Three (3) Channel Digital Recorder<br>- Consistency<br>- Temperature<br>- Pressure- (work with Automatic Pressure Control) | Required      |