

EVALUATION REPORT
(As Per Rule 35 of PPRA Rules, 2004)

1.	Name of Procuring Agency:	Oil & Gas Development Company Limited.			
2.	Method of Procurement:	Press Tender (Single Stage – Two Envelope)			
3.	Title of Procurement:	Online Moisture Analyzer for Sale Gas Metering			
4.	Tender Inquiry No.	PROC-FE/CB/PE&FD/TOGH-4904/2020			
5.	PPRA Ref. No. (TSE):	TSE. No. not available at PPRA Website as it is available only for 10 days from the closing date.			
6.	Date & Time of Bid Closing:	26-01-2021– 10:00 Hrs			
7.	Date & Time of Bid Opening:	26-01-2021– 10:30 Hrs			
8.	No. of Bids Received:	Four (04)			
9.	Criteria for Bid Evaluation:	Main Item/Equipment Wise			
10.	Detail of Bid (s) Evaluation:	As below			
Name of Bidder	Marks		Cost (Rs.)		Basis for Rejection/Acceptance of the Bids (Clause # 35 of PPRA)
	Technical (If applicable)	Financial (If applicable)	Quoted Cost	Evaluated Cost	
M/s Michell Instruments Limited, UK	N/A	N/A	GBP 31,118/- (equivalent to USD 42,454.90)	GBP 31,118/- (equivalent to USD 42,454.90)	Technically Responsive & Financially 1 st Lowest as per Clause No. 38 of PPRA
M/s Ametek Process Instruments, USA	N/A	N/A	USD 57,852.94	USD 57,852.94	Technically Responsive & Financially 2 nd Lowest as per Clause No. 38 of PPRA
M/s Texas Petroleum Supplies Corporation, USA	N/A	N/A	N/A	N/A	Technically Non-Responsive due to following reasons: i. Built-in Moisture Analyzer Generator is not available in the quoted model. This is considered as material specifications deviation as per tender clause 22 of Annexure-Z of Tender specification. ii. Sample Probe is not quoted as required vide Clause 32 of Annexure-Z of Tender Specification.
M/s Solutronix Trading FZE, UAE	N/A	N/A	N/A	N/A	Technically Non-Responsive due to following reason: i. Built-in Moisture Analyzer Generator is not available in the quoted model. This is considered as material specifications deviation as per tender clause 22 of Annexure-Z of Tender specification.
Lowest Evaluated Bidder: M/s Michell Instruments Limited, UK					