



Technical Evaluation Report

For the Procurement of "Spectrophotometer" against
TE # (PROC-LE/PT/EXP/17466/18)

Sr. No	Name of Bidder	Technical Responsive (Yes / No)	Reason of Rejection
1	Samteck Energy Tools	Yes	N/A
2	SAT International	No	The bidder is not agreed to withdraw price fluctuation clause
3.1	National Engineering Service Trading and consultancy company (NESTAC) (Option A, Spectrophotometer 6850)	No	The quoted specifications of the instrument are not as per Annexure A of SOR. The quoted instrument does not offer specifically designed & built-in methods for water analysis. It also does not offer reagent kits for water analysis.
3.2	National Engineering Service Trading and consultancy company (NESTAC) (Option B, Hach DR-6000)	Yes	N/A
4	Continental Trading	No	The quoted specifications of the instrument are not as per Annexure A of SOR. The required wavelength range was 190-1100 nm, whereas the quoted instrument offers 320-1100 nm. The parameters namely accuracy and band width are also different from the those specified in SOR. The bidder is not agreed to deliver (As per letter Ref: CTC-548/S dated. 04/12/18) within time specified in SOR.


Muhammad Adnan Saqlain
Assistant Lab Chemist


Rafiq Ahmad
Sr. Tech. Officer Lab