

**TECHNICAL EVALUATION REPORT**

**TENDER NO. PROC-FE/CB/PROD-3069/2017**

**FOR CORIOLIS MASS FLOW METERS**

<b>Sr. No</b>	<b>Bidder Name</b>	<b>Technically Responsive (Yes/No)</b>	<b>Reason for Rejection</b>
1	<b>M/s Samtek International Inc., USA</b> C/o M/s Samtek International, Islamabad	Yes	-
2	<b>M/s Solutronix Trading FZE, UAE .</b> C/o M/s Solutions Engineering Pvt. Ltd., Lahore .	Yes	-
3	<b>M/s Intech Automazione SRL, Italy .</b> C/o M/s Dynamic Engineering and Automation, Islamabad.	Yes	
4	<b>M/s FMA System Limited, UK.</b> C/O M/s Atlas Steel & Engineering Co., Karachi.	No	i. Quoted model does not comply the custody transfer approvals of PTB / METAS / BEV / MID / NTEP / MC. ii. Quoted model does not comply the Standards/Approvals/Test, Certificate: EN1289-2x (PT) + EN12517-1-1 (RT) meas. tube (PT) +process connection (RT) welding seam, inspection certificate.
5	<b>M/s Shengli Oil field Shengli Petroleum Inst: Factory,China .</b> C/o M/s Speedy Automation., Islamabad	No	i. Quoted model does not comply the custody transfer approvals. This is considered as material deviation w.r.t to tender specification of ANNEX Z