

TECHNICAL EVALUATION REPORT

TENDER NO. PROC-FE/CB/PROD-4379/2019

FOR INSTRUMENTATION MATERIAL & EXPLOSION PROOF LIGHTS

Sr. No	Bidder Name	Technically Responsive (Yes/No)	Reason for Rejection
1	M/s WIKA, UAE. C/o M/s Maxwell Automation, Lahore.	No (for group "B")	1. Quoted model of Digital Temperature Gauges (DTGs) doesn't possess the feature of explosion proof, which deviates the OGDCL requirement of Explosion Proof Digital Temperature Gauges mentioned at Sr. No 1 & 13 of Annexure X of Digital Temperature Gauge data sheet specifications of tender document. 2. Accuracy of quoted model (WIKA TR75) of DTGs was 0.5% which deviates the OGDCL requirement of 0.1% mentioned at Sr. No 6 of Annexure X of Digital Temperature Gauge data sheet specifications of tender document.
2	M/s Zain Enterprises, FZE, UAE. C/o Brittle Engineering Company, Karachi.	Yes (for group "C")	-
3	M/s INDENTORS International, USA. C/o M/s INDENTORS International, Islamabad.	Yes (for group "C")	-
4	M/s FLOW Trading L.L.C, UAE. C/o M/s MAS Associates Pvt, Ltd., Lahore.	Yes (for group "C")	-
5	M/s Warom Technology Inc. Company, China. C/o M/s Maxwell Automation, Lahore.	Yes (for group "C")	-

Note: No bid has been received against Group "A".