

**Technical Evaluation Performa**  
Gastec Tubes (T.E No. PROC-FE/CB/P&P4383/2019)

S/No.	Bidder/Manufacturer	Technically responsive (Y/N)	Remarks
1	<b>M/s A1-Cbiss Ltd., UK</b> (M/s Nescon International, Islamabad)	Yes	~
2	<b>M/s A1-Cbiss Ltd., UK</b> (M/s Speedy Automation, Islamabad)	Yes	~
	<b>M/s ABCOP International Inc., USA</b> (M/s Chem-Tech International, Islamabad)	No	Did not comply tender requirement. <ul style="list-style-type: none"><li>• Quoted tubes against serial number 7-11, 16, 18 and 24-29 of the SOR does not follow required measuring ranges.</li><li>• Local agent failed to provide/confirm compatibility of quoted tubes with existing in use Gastec pumps.</li></ul>