

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

Item No.	Sr. No.	Name of Bidder	Status (Technically Responsive/ Non- Responsive)	Remarks (Reasons of Rejection)
1	1	Enersol International Ltd, Canada (ESP Safety Inc., USA)	Technically Responsive	-
	2	Al Murabit Co, Dubai, U.A.E (HONEYWELL LIFE SAFETY DISTRIBUTION GmbH)	Technically Non-Responsive	Non-compliance with Clause 16, 17, 18, 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period. Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life. Quoted H2S Sensor's operating temperature rating is -20°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±5ppm while required accuracy is ±3% of full scale (i.e. ±3ppm at 100 PPM range)
	3	Nuricon Petro services Fze., U.A.E (Nanjing Aiyi Technologies Co., Ltd., China)	Technically Non-Responsive	Non-compliance with Clause 16, 18 & 19 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 6 months shelf life of sensor instead of required 1 year shelf life.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT

RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

				<ul style="list-style-type: none">Quoted H2S Sensor's operating temperature rating is -40°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. Non-compliance with Clause 30 of TOR Annexure-Z Transmitter Specifications: <ul style="list-style-type: none">Quoted Aluminum housing for detector instead of required SS316.
	4	IES FZE, U.A.E (Teledyne Detcon, USA)	Technically Non-Responsive	Non-compliance with Clause 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C.Accuracy of quoted H2S Sensor is ±10% while required accuracy is ±3% of full scale.
	5	Solutronix Trading FZE, U.A.E (Gasensor Technology Pte Ltd., Singapore)	Technically Responsive	-
	6	TTM General Trading LLC, U.A.E (Riken Keiki Co Ltd, Japan)	Technically Non-Responsive	Non-compliance with Clause 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Accuracy of quoted H2S Sensor is ±5% of full scale while required accuracy is ±3% of full scale.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT

RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	7	Dominar engineers LLC, FZ, U.A.E (Hanwei Electronics group Corporation, China)	Technically Non-Responsive	Non-compliance with Clause 19 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C.
	8	Descon Engineering FZE, Sharjah, U.A.E (MSA Middle East FZE)	Technically Responsive	-
	9	Drager Safety AG & Co. KGaA, Lubeck (Drager Safety AG & Co. KGaA, Lubeck)	Technically Responsive	-
	10	Rah Enterprises, Sharjah, U.A.E (RC SYSTE MS INC, USA)	Technically Non-Responsive	Non-compliance with Clause 17, 18, 19, 20 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 18 months warrantee period of sensor instead of required 2 years' warrantee period.Quoted 6 months shelf life of sensor instead of required 1 year shelf life.Quoted H2S Sensor's operating temperature rating is -30°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C.Response time T (90) of quoted H2S Sensor is <55 sec while required Response time T (90) is <50 sec.Accuracy of quoted H2S Sensor is ±4% while required accuracy is ±3% of full scale.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	11	Al Masood Oil Industry Supplies & Services Co. U.A.E (Crowcon, UK)	Technically Non-Responsive	Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted sensor's shelf life is 6 months instead of required 1 year shelf life. Incomplete/unsatisfactory response against the clarifications requested via Ariba message board dated 21-07-2025. Despite multiple reminders sent on 29-07-2025, 01-08-2025, and 04-08-2025, no further response or required supporting documents/evidence have been received to date.
	12	Ruien, Lane, Xing Pu Road, Suzhou Industrial Park, Jiangsu, P.R. China (Honeywell Automation and Control (China) Co., Ltd)	Technically Non-Responsive	Non-compliance with Clause 16, 17, 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 2 years Sensor's operating life instead of required 3 years' operating life.Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period.Quoted H2S Sensor's operating temperature rating is -20°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C.Accuracy of quoted H2S Sensor is ±5ppm while required accuracy is ±3% of full scale (i.e. ±3ppm at 100 PPM range)
	13	Avanceon FZE, Dubai, U.A.E (M/s Honeywell Head Quarters, Switzerland)	Technically Non-Responsive	Non-compliance with Clause 16, 17, 18, 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 2 years Sensor's operating life instead of required 3 years' operating life.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

				<ul style="list-style-type: none"> Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period. Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life. Quoted H2S Sensor's operating temperature rating is -20°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±5ppm while required accuracy is ±3% of full scale (i.e. ±3ppm at 100 PPM range)
	14	Morrells International General Trading, Bahrain (CROWCON UK)	Technically Non-Responsive	<p>Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none"> Quoted 6 months shelf life of sensor instead of required 1 year shelf life. <p>Non-compliance with Clause 36 of TOR Annexure-Z Calibration Gas Cylinders:</p> <ul style="list-style-type: none"> Quoted calibration Gas cylinders have 12 months shelf life instead of required 24 months shelf life
2	1	Al Murabit Co, Dubai, U.A.E (HONEYWELL LIFE SAFETY DISTRIBUTION GmbH)	Technically Non-Responsive	<p>Non-compliance with Clause 16, 17, 18, 19 & 22 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none"> Quoted 2 years Sensor's operating life instead of required 3 years' operating life. Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period. Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

				<ul style="list-style-type: none"> Quoted H2S Sensor's operating temperature rating is -20°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±5ppm while required accuracy is ±3% of full scale (i.e. ±3ppm at 100 PPM range)
	2	Nuricon Petroservices Fze., U.A.E (Nanjing Aiyi Technologies Co., Ltd., China)	Technically Non-Responsive	<p>Non-compliance with Clause 16, 18 & 19 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 6 months shelf life of sensor instead of required 1 year shelf life. Quoted H2S Sensor's operating temperature rating is -40°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. <p>Non-compliance with Clause 30 of TOR Annexure-Z Transmitter Specifications:</p> <ul style="list-style-type: none"> Quoted Aluminum housing for detector instead of required SS316.
	3	IES FZE, U.A.E (Teledyne Detcon, USA)	Technically Non-Responsive	<p>Non-compliance with Clause 19 & 22 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none"> Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±10% while required accuracy is ±3% of full scale.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT

RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	4	Solutronix Trading FZE, U.A.E (Gasensor Technology Pte Ltd., Singapore)	Technically Responsive	-
	5	Dominar engineers LLC, FZ, U.A.E (Hanwei Electronics group Corporation, China)	Technically Non-Responsive	Non-compliance with Clause 19 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C.
	6	Descon Engineering FZE, Sharjah, U.A.E (MSA Middle East FZE)	Technically Responsive	-
	7	Rah Enterprises, Sharjah, U.A.E (RC SYSTE MS INC, USA)	Technically Non-Responsive	Non-compliance with Clause 17, 18, 19, 20 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 18 months warrantee period of sensor instead of required 2 years' warrantee period.Quoted 6 months shelf life of sensor instead of required 1 year shelf life.Quoted H2S Sensor's operating temperature rating is -30°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C.Response time T (90) of quoted H2S Sensor is <55 sec while required Response time T (90) is <50 sec.Accuracy of quoted H2S Sensor is ±4% while required accuracy is ±3% of full scale.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	8	BAOJI MENGTAI PETROLEUM MACHINERY CO. LTD, china (HANWEI ELECTRONICS GROUP CORPORATION, China)	Technically Non-Responsive	Non-compliance with Clause 19 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±5% while required accuracy is ±3% of full scale.
	9	Al Masood Oil Industry Supplies & Services Co. U.A.E (Crowcon, UK)	Technically Non-Responsive	Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted sensor's shelf life is 6 months instead of required 1 year shelf life. Incomplete/unsatisfactory response against the clarifications requested via Ariba message board dated 21-07-2025. Despite multiple reminders sent on 29-07-2025, 01-08-2025, and 04-08-2025, no further response or required supporting documents/evidence have been received to date.
	10	Ruien, Lane, Xing Pu Road, Suzhou Industrial Park, Jiangsu, P.R. China (Honeywell Automation and Control (China) Co., Ltd)	Technically Non-Responsive	Non-compliance with Clause 16, 17, 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period. Quoted H2S Sensor's operating temperature rating is -20°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT

RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

				<ul style="list-style-type: none">Accuracy of quoted H2S Sensor is ± 5ppm while required accuracy is $\pm 3\%$ of full scale (i.e. ± 3ppm at 100 PPM range)
	11	Avanceon FZE, Dubai, U.A.E (M/s Honeywell Head Quarters, Switzerland)	Technically Non- Responsive	Non-compliance with Clause 16, 17, 18, 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 2 years Sensor operating life instead of required 3 years' operating life.Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period.Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life.Quoted H2S Sensor's operating temperature rating is $-20^{\circ}\text{C} \sim +55^{\circ}\text{C}$ whereas required operating temperature rating is $-10^{\circ}\text{C} \sim +60^{\circ}\text{C}$.Accuracy of quoted H2S Sensor is ± 5ppm while required accuracy is $\pm 3\%$ of full scale (i.e. ± 3ppm at 100 PPM range)
	12	Morrells International General Trading, Bahrain (CROWCON UK)	Technically Non- Responsive	Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 6 months shelf life of sensor instead of required 1 year shelf life. Non-compliance with Clause 36 of TOR Annexure-Z Calibration Gas Cylinders: <ul style="list-style-type: none">Quoted calibration Gas cylinders have 12 months shelf life instead of required 24 months shelf life.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

3	1	Enersol International Ltd, Canada (ESP Safety Inc., USA)	Technically Responsive	-
	2	Al Murabit Co, Dubai, U.A.E (HONEYWELL LIFE SAFETY DISTRIBUTION GmbH)	Technically Non-Responsive	Non-compliance with Clause 16, 17 & 18 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period. Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life.
	3	Nuricon Petroservices Fze., U.A.E (Nanjing Aiyi Technologies Co., Ltd., China)	Technically Non-Responsive	Non-compliance with Clause 16, 18 & 19 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 2 years Sensor operating life instead of required 3 years' operating life. Quoted 6 months shelf life of sensor instead of required 1 year shelf life. Quoted H2S Sensor's operating temperature rating is -40°C ~ +55°C whereas required operating temperature rating is -10°C ~ +60°C. Non-compliance with Clause 30 of TOR Annexure-Z Transmitter Specifications: <ul style="list-style-type: none"> Quoted Aluminum housing for detector instead of required SS316.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	4	IES FZE, U.A.E (Teledyne Detcon, USA)	Technically Non-Responsive	Non-compliance with Clause 19 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted H2S Sensor's operating temperature rating is -20°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C. Accuracy of quoted H2S Sensor is ±10% while required accuracy is ±3% of full scale.
	5	Solutronix Trading FZE, U.A.E (Gasensor Technology Pte Ltd., Singapore)	Technically Responsive	-
	6	Descon Engineering FZE, Sharjah, U.A.E (MSA Middle East FZE)	Technically Responsive	-
	7	Rah Enterprises, Sharjah, U.A.E (RC SYSTEMS INC, USA)	Technically Non-Responsive	Non-compliance with Clause 17, 18, 19, 20 & 22 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none"> Quoted 18 months warrantee period of sensor instead of required 2 years' warrantee period. Quoted 6 months shelf life of sensor instead of required 1 year shelf life. Quoted H2S Sensor's operating temperature rating is -30°C ~ +50°C whereas required operating temperature rating is -10°C ~ +60°C. Response time T (90) of quoted H2S Sensor is <55 sec while required Response time T (90) is <50 sec. Accuracy of quoted H2S Sensor is ±4% while required accuracy is ±3% of full scale.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	8	Al Masood Oil Industry Supplies & Services Co. U.A.E (Crowcon, UK)	Technically Non-Responsive	<p>Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none">Quoted sensor's shelf life is 6 months instead of required 1 year shelf life. <p>The quoted H₂S Gas Detector does not support dual sensor capability, it is designed to operate with only one sensor at a time.</p> <p>Incomplete/unsatisfactory response against the clarifications requested via Ariba message board dated 21-07-2025. Despite multiple reminders sent on 29-07-2025, 01-08-2025, and 04-08-2025, no further response or required supporting documents/evidence have been received to date.</p>
	9	Ruien, Lane, Xing Pu Road, Suzhou Industrial Park, Jiangsu, P.R. China (MSA, USA)	Technically Responsive	-
	10	Avanceon FZE, Dubai, U.A.E (M/s Honeywell Head Quarter s, Switzerland)	Technically Non-Responsive	<p>Non-compliance with Clause 16, 17, 18, 19 & 22 of TOR Annexure-Z Sensor Specifications:</p> <ul style="list-style-type: none">Quoted 2 years Sensor operating life instead of required 3 years' operating life.Quoted 1 year warrantee period of sensor instead of required 2 years' warrantee period.Quoted 6-9 months shelf life of sensor instead of required 1 year shelf life.

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.

TECHNICAL EVALUATION REPORT
RFP NO.: OGDCL -SCM-FD-CB-PEnFD--283920262-2025

DESCRIPTION OF THE RFP: H2S GAS DETECTORS

TECHNICAL BIDS OPENING DATE: 26-05-2025

Evaluation Criteria: Item Wise

	11	Morrells International General Trading, Bahrain (CROWCON UK)	Technically Non- Responsive	Non-compliance with Clause 18 of TOR Annexure-Z Sensor Specifications: <ul style="list-style-type: none">Quoted 6 months shelf life of sensor instead of required 1 year shelf life. Non-compliance with Clause 36 of TOR Annexure-Z Calibration Gas Cylinders: <ul style="list-style-type: none">Quoted calibration Gas cylinders have 12 months shelf life instead of required 24 months shelf life The quoted H₂S Gas Detector does not support dual sensor capability, it is designed to operate with only one sensor at a time.
--	----	---	--	--

Note:

- Bidder may submit grievance within Seven (07) days of announcement of Technical Evaluation Report.
- The grievance if any shall be sent to bgrc@ogdcl.com as per format available on OGDCL website.