

Technical Evaluation Report
Detector Tubes
(Proc-FE/CB/P&Lab-5283/2022)

S/N o.	Manufacturer/Bidder/Local Agent	Technically responsive Y/N	Remarks
1	M/s A1-CBISS/Gastec Corporation M/s Indman International Inc. USA M/s Acumen International, Islamabad.	Yes	~
2	M/s Grand Line, Canada M/s Grand Line, Canada M/s Nescon International, Islamabad	Yes	~
3	M/s Dragger Safety AG& Co, KGaA M/s Dragger Safety AG& Co, KGaA M/s ABGURT (Pvt.)Ltd. Karachi	No	Does not comply with tender requirement, <ul style="list-style-type: none"> Does not comply with tender specification, Quoted detector tubes deviate from indented ranges and as per tender clause 1.2.7 are not compatible with the Gastec pump.

15/6/22
YAHYA TAJVEED
 Manager (P&P)-CA/T&SS
 Ext 3674

16-6-22
Hajra Khatoon
 Sr. Chemist (P&P)
 OGDCL H.O, Ex. 3644