

**TECHNICAL EVALUATION REPORT**  
**MUD LABORATORY EQUIPMENT**

**CASE NO: PROC-FE/CB/WS-5052/21**

**MUD LABORATORY EQUIPMENT (GROUP A)**

<b>S NO.</b>	<b>BIDDER'S NAME</b>	<b>LOCAL AGENT</b>	<b>TECHNICALLY RESPONSIVE (YES/NO)</b>	<b>REASONS FOR REJECTION</b>
1.	M/s Halliburton Worldwide Limited, Islamabad.	M/s Halliburton Worldwide Limited, Islamabad.	YES	
2.	M/S Woodhouse International Fze.	M/s Redox (Private) LTD, Islamabad.	YES	

**TECHNICAL EVALUATION REPORT**  
**MUD LABORATORY EQUIPMENT**

CASE NO: PROC-FE/CB/WS-5052/21

**MUD LABORATORY EQUIPMENT (GROUP B)**

S NO.	BIDDER'S NAME	LOCAL AGENT	TECHNICALLY RESPONSIVE (YES/NO)	REASONS FOR REJECTION
1.	M/s Halliburton Worldwide Limited, Islamabad.	M/s Halliburton Worldwide Limited, Islamabad.	NO	<p>Three out of five items in group B does not compliance with tender document specification (Annexure-A1).</p> <p><b>1. Hot Plate with magnetic Stirrer:</b> THERMOLINE (205742) USA does not meet required operating temperature specification i.e 480°C and quoted model does not have provision of magnetic stirring as only hot plate is quoted.</p> <p><b>2. Water Distillation apparatus:</b> No bid submitted</p> <p><b>3. Heating Oven:</b> Bidder quoted <b>Roller</b> Heating Oven while <b>Static</b> Heating Oven is published in tender document.</p>
2.	M/s WoodHouse international FZE	M/s Redox (Private) LTD, Islamabad.	NO	<p>I. During technical evaluation stage upon clarification, the local agent (M/s Redox PVT Ltd.) declared himself bidder for Group B.</p> <p>II. M/s WoodHouse international FZE being bidder No. 2 did not participated in Group B (Annexure-E "Data Summary Sheet" not submitted).</p> <p>III. M/s Redox PVT Ltd submitted Annexure-E "Data Summary Sheet for Group B" but without other required documents i.e ISO, Manufacturer authority letters and Five (05) year sale record and offered to provide group B items through local distributor (M/s Science Emporium, Lahore), which is not acceptable as per OGDCL procurement rules.</p>

*S. Ullah*  
30/8/21  
**AMTULLAH**  
Mud Engine  
Ext: 2306